



Application parameters for:

HITACHI 704

HITACHI 704 پارامترهای کیت ALBUMIN برای دستگاه
با روش BCG

TEST	ALB
ASSAY CODE	1 POINT - 32 - 0
SAMPLE VOLUME (μl)	3
R 1 VOLUME (μl)	300 - 50 - NO
R 2 VOLUME (μl)	0 - 20 - NO
WAVELENGTH (nm)	700 - 546
CALIBR. METHOD	1 - 0 - 0
STD.1 CONC.-POS.	0.0 - 1
STD.2 CONC.-POS.	* - *
STD.3 CONC.-POS.	0 - 0
STD.4 CONC.-POS.	0 - 0
STD.5 CONC.-POS.	0 - 0
STD.6 CONC.-POS.	0 - 0
SD LIMIT	0.1
DUPLICATE LIMIT	200
SENSITIVITY LIMIT	0
ABS.LIMIT(INC/DEC)	0 - 0
PROZONE LIMIT	0 - 0
EXPECTED VALUE	3.5 - 5.2
UNIT	Mg/dl
INSTRUMENT FACTOR	1.00

پارامترهای کیت ALKALINE PHOSPHATASE برای دستگاه

HITACHI 704

با روش DGKC

TEST	ALP
ASSAY CODE	RATE A - 19 - 32
SAMPLE VOLUME (μl)	6
R 1 VOLUME (μl)	240 - 50 - NO
R 2 VOLUME (μl)	60 - 20 - NO
WAVELENGTH (nm)	660 - 415
CALIBR. METHOD	1 - 0 - 0
STD.1 CONC.-POS.	0 - 1
STD.2 CONC.-POS.	* - *
STD.3 CONC.-POS.	0 - 0
STD.4 CONC.-POS.	0 - 0
STD.5 CONC.-POS.	0 - 0
STD.6 CONC.-POS.	0 - 0
SD LIMIT	0.1
DUPLICATE LIMIT	200
SENSITIVITY LIMIT	0
ABS.LIMIT(INC/DEC)	13000 - 0
PROZONE LIMIT	0 - 0
EXPECTED VALUE	64 - 306
UNIT	U/L
INSTRUMENT FACTOR	1.00

HITACHI 704 پارامترهای کیت ALT (SGPT) برای دستگاه
با روش IFCC

TEST	ALT
ASSAY CODE	RATE A - 19 - 28
SAMPLE VOLUME (μl)	20
R 1 VOLUME (μl)	240 - 50 - NO
R 2 VOLUME (μl)	60 - 20 - NO
WAVELENGTH (nm)	415 - 340
CALIBR. METHOD	1 - 0 - 0
STD.1 CONC.-POS.	0.0 - 1
STD.2 CONC.-POS.	* - *
STD.3 CONC.-POS.	0 - 0
STD.4 CONC.-POS.	0 - 0
STD.5 CONC.-POS.	0 - 0
STD.6 CONC.-POS.	0 - 0
SD LIMIT	0.1
DUPLICATE LIMIT	200
SENSITIVITY LIMIT	0
ABS.LIMIT(INC/DEC)	8000 - 1
PROZONE LIMIT	0 - 0
EXPECTED VALUE	0 - 41
UNIT	U/L
INSTRUMENT FACTOR	1.00

HITACHI 704 پارامترهای کیت AMYLASE برای دستگاه
با روش EPS-G7 (DEA)

TEST	AMY
ASSAY CODE	RATE A - 6 - 18
SAMPLE VOLUME (μl)	6
R 1 VOLUME (μl)	300 - 50 - NO
R 2 VOLUME (μl)	0 - 20 - NO
WAVELENGTH (nm)	660 - 405
CALIBR. METHOD	1- 0 - 0
STD.1 CONC.-POS.	0- 1
STD.2 CONC.-POS.	* - *
STD.3 CONC.-POS.	0- 0
STD.4 CONC.-POS.	0- 0
STD.5 CONC.-POS.	0- 0
STD.6 CONC.-POS.	0- 0
SD LIMIT	0.1
DUPLICATE LIMIT	200
SENSITIVITY LIMIT	0
ABS.LIMIT(INC/DEC)	10000 - 0
PROZONE LIMIT	0- 0
EXPECTED VALUE	0- 100
UNIT	U/L
INSTRUMENT FACTOR	1.00

پارامترهای کیت AST (SGOT) برای دستگاه **HITACHI 704**
با روش IFCC

TEST	AST
ASSAY CODE	RATE A - 19 - 28
SAMPLE VOLUME (μl)	20
R 1 VOLUME (μl)	240 - 50 - NO
R 2 VOLUME (μl)	60 - 20 - NO
WAVELENGTH (nm)	415 - 340
CALIBR. METHOD	1 - 0 - 0
STD.1 CONC.-POS.	0.0 - 1
STD.2 CONC.-POS.	* - *
STD.3 CONC.-POS.	0 - 0
STD.4 CONC.-POS.	0 - 0
STD.5 CONC.-POS.	0 - 0
STD.6 CONC.-POS.	0 - 0
SD LIMIT	0.1
DUPLICATE LIMIT	200
SENSITIVITY LIMIT	0
ABS.LIMIT(INC/DEC)	6000 - 1
PROZONE LIMIT	0 - 0
EXPECTED VALUE	0 - 37
UNIT	U/L
INSTRUMENT FACTOR	1.00

HITACHI پارامتر های کیت BILIRUBIN DIRECT برای دستگاه

704

با روش DCA

TEST	BILD
ASSAY CODE	2 POINT - 15 - 32
SAMPLE VOLUME (μl)	20
R 1 VOLUME (μl)	300 - 50 - NO
R 2 VOLUME (μl)	75 - 20 - NO
WAVELENGTH (nm)	660 - 546
CALIBR. METHOD	1 - 0 - 0
STD.1 CONC.-POS.	0.00 - 1
STD.2 CONC.-POS.	* - *
STD.3 CONC.-POS.	0 - 0
STD.4 CONC.-POS.	0 - 0
STD.5 CONC.-POS.	0 - 0
STD.6 CONC.-POS.	0 - 0
SD LIMIT	0.1
DUPLICATE LIMIT	400
SENSITIVITY LIMIT	0
ABS.LIMIT(INC/DEC)	0 - 0
PROZONE LIMIT	0 - 0
EXPECTED VALUE	0.00 - 0.30
UNIT	Mg/dl
INSTRUMENT FACTOR	1.00

HITACHI 704 پارامترهای کیت BILIRUBIN TOTAL برای دستگاه
با روش DCA

TEST	BILT
ASSAY CODE	2 POINT - 15 - 32
SAMPLE VOLUME (μl)	15
R 1 VOLUME (μl)	300 - 50 - NO
R 2 VOLUME (μl)	75 - 20 - NO
WAVELENGTH (nm)	660 - 546
CALIBR. METHOD	1 - 0 - 0
STD.1 CONC.-POS.	0.0 - 1
STD.2 CONC.-POS.	* - *
STD.3 CONC.-POS.	0 - 0
STD.4 CONC.-POS.	0 - 0
STD.5 CONC.-POS.	0 - 0
STD.6 CONC.-POS.	0 - 0
SD LIMIT	0.1
DUPLICATE LIMIT	400
SENSITIVITY LIMIT	0
ABS.LIMIT(INC/DEC)	0 - 0
PROZONE LIMIT	0 - 0
EXPECTED VALUE	0.1 - 1.2
UNIT	Mg/dl
INSTRUMENT FACTOR	1.00

HITACHI 704 پارامترهای کیت CALCIUM برای دستگاه
با روش ARSENASO

TEST	CA
ASSAY CODE	1 POINT - 15 - 0
SAMPLE VOLUME (μl)	3
R 1 VOLUME (μl)	300 - 50 - NO
R 2 VOLUME (μl)	0 - 20 - NO
WAVELENGTH (nm)	0 - 660
CALIBR. METHOD	1 - 0 - 0
STD.1 CONC.-POS.	0.0 - 1
STD.2 CONC.-POS.	* - *
STD.3 CONC.-POS.	0 - 0
STD.4 CONC.-POS.	0 - 0
STD.5 CONC.-POS.	0 - 0
STD.6 CONC.-POS.	0 - 0
SD LIMIT	0.1
DUPLICATE LIMIT	500
SENSITIVITY LIMIT	0
ABS.LIMIT(INC/DEC)	0 - 0
PROZONE LIMIT	0 - 0
EXPECTED VALUE	8.6 - 10.3
UNIT	Mg/dl
INSTRUMENT FACTOR	1.00

HITACHI 704 پارامتر های کیت CHOLESTEROL برای دستگاه
با روش GOD

TEST	CHOL
ASSAY CODE	1 POINT- 32 - 0
SAMPLE VOLUME (μl)	3
R 1 VOLUME (μl)	300 - 50 - NO
R 2 VOLUME (μl)	0- 20 - NO
WAVELENGTH (nm)	700 - 505
CALIBR. METHOD	1- 0 - 0
STD.1 CONC.-POS.	0- 1
STD.2 CONC.-POS.	* - *
STD.3 CONC.-POS.	0 - 0
STD.4 CONC.-POS.	0 - 0
STD.5 CONC.-POS.	0 - 0
STD.6 CONC.-POS.	0 - 0
SD LIMIT	0.1
DUPLICATE LIMIT	400
SENSITIVITY LIMIT	0
ABS.LIMIT(INC/DEC)	0 - 0
PROZONE LIMIT	0 - 0
EXPECTED VALUE	0- 200
UNIT	Mg/dl
INSTRUMENT FACTOR	1.00

HITACHI 704 پارامترهای کیت CK-MB برای دستگاه

TEST	CKMB
ASSAY CODE	RATE A - 22 - 32
SAMPLE VOLUME (μl)	12
R 1 VOLUME (μl)	240 - 50 - NO
R 2 VOLUME (μl)	60 - 20 - NO
WAVELENGTH (nm)	415 - 340
CALIBR. METHOD	1 - 0 - 0
STD.1 CONC.-POS.	0 - 1
STD.2 CONC.-POS.	* - *
STD.3 CONC.-POS.	0 - 0
STD.4 CONC.-POS.	0 - 0
STD.5 CONC.-POS.	0 - 0
STD.6 CONC.-POS.	0 - 0
SD LIMIT	0.1
DUPLICATE LIMIT	200
SENSITIVITY LIMIT	0
ABS.LIMIT(INC/DEC)	9000 - 0
PROZONE LIMIT	0 - 0
EXPECTED VALUE	0 - 24
UNIT	U/L
INSTRUMENT FACTOR	1.00

HITACHI 704 پارامترهای کیت CPK برای دستگاه
با روش IFCC/DGKC

TEST	CPK
ASSAY CODE	RATE A - 22 - 32
SAMPLE VOLUME (μl)	1 2
R 1 VOLUME (μl)	240 - 50 - NO
R 2 VOLUME (μl)	60 - 20 - NO
WAVELENGTH (nm)	415 - 340
CALIBR. METHOD	1 - 0 - 0
STD.1 CONC.-POS.	0 - 1
STD.2 CONC.-POS.	* - *
STD.3 CONC.-POS.	0 - 0
STD.4 CONC.-POS.	0 - 0
STD.5 CONC.-POS.	0 - 0
STD.6 CONC.-POS.	0 - 0
SD LIMIT	0.1
DUPLICATE LIMIT	200
SENSITIVITY LIMIT	0
ABS.LIMIT(INC/DEC)	9000 - 0
PROZONE LIMIT	0 - 0
EXPECTED VALUE	24 - 195
UNIT	U/L
INSTRUMENT FACTOR	1.00

HITACHI 704 پارامترهای کیت CREATININE برای دستگاه
با روش JUFFE

TEST	CREAT
ASSAY CODE	2 POINT - 18 - 21
SAMPLE VOLUME (μl)	2 0
R 1 VOLUME (μl)	300 - 50 - NO
R 2 VOLUME (μl)	75 - 20 - NO
WAVELENGTH (nm)	660 - 505
CALIBR. METHOD	1 - 0 - 0
STD.1 CONC.-POS.	0 - 1
STD.2 CONC.-POS.	* - *
STD.3 CONC.-POS.	0 - 0
STD.4 CONC.-POS.	0 - 0
STD.5 CONC.-POS.	0 - 0
STD.6 CONC.-POS.	0 - 0
SD LIMIT	0.1
DUPLICATE LIMIT	200
SENSITIVITY LIMIT	0
ABS.LIMIT(INC/DEC)	4500 - 0
PROZONE LIMIT	0 - 0
EXPECTED VALUE	0.7 - 1.4
UNIT	Mg/dl
INSTRUMENT FACTOR	1.00

HITACHI 704 پارامترهای کیت GLUCOSE برای دستگاه
با روش GOD/POD

TEST	GLU
ASSAY CODE	1 POINT - 32 - 0
SAMPLE VOLUME (μl)	3
R 1 VOLUME (μl)	300 - 50 - NO
R 2 VOLUME (μl)	0 - 20 - NO
WAVELENGTH (nm)	700 - 505
CALIBR. METHOD	1 - 0 - 0
STD.1 CONC.-POS.	0.0 - 1
STD.2 CONC.-POS.	* - *
STD.3 CONC.-POS.	0 - 0
STD.4 CONC.-POS.	0 - 0
STD.5 CONC.-POS.	0 - 0
STD.6 CONC.-POS.	0 - 0
SD LIMIT	0.1
DUPLICATE LIMIT	400
SENSITIVITY LIMIT	0
ABS.LIMIT(INC/DEC)	0 - 0
PROZONE LIMIT	0 - 0
EXPECTED VALUE	70 - 115
UNIT	Mg/dl
INSTRUMENT FACTOR	1.00

HITACHI 704 پارامترهای کیت Hb A1c برای دستگاه

TEST	HbA1c
ASSAY CODE	1 POINT- 32 - 0
SAMPLE VOLUME (μl)	5
R 1 VOLUME (μl)	185 - 20 - NO
R 2 VOLUME (μl)	60 - 20 - NO
WAVELENGTH (nm)	0 - 660
CALIBR. METHOD	3 - 4 - 5
STD.1 CONC.-POS.	0.0 - 1 Nacl
STD.2 CONC.-POS.	* - *
STD.3 CONC.-POS.	* - *
STD.4 CONC.-POS.	* - *
STD.5 CONC.-POS.	* - *
STD.6 CONC.-POS.	0 - 0
SD LIMIT	999
DUPLICATE LIMIT	1000
SENSITIVITY LIMIT	0
ABS.LIMIT(INC/DEC)	32000 - 0
PROZONE LIMIT	0 - 0
EXPECTED VALUE	3.0 - 6.0
UNIT	%
INSTRUMENT FACTOR	1.0

HITACHI 704 پارامتر های کیت HDL برای دستگاه
با روش DIRECT

TEST	HDL-C
ASSAY CODE	2 POINT - 18 - 32
SAMPLE VOLUME (μl)	3
R 1 VOLUME (μl)	270 - 50 - NO
R 2 VOLUME (μl)	90 - 20 - NO
WAVELENGTH (nm)	700 - 600
CALIBR. METHOD	1- 0 - 0
STD.1 CONC.-POS.	0- 1
STD.2 CONC.-POS.	* - *
STD.3 CONC.-POS.	0 - 0
STD.4 CONC.-POS.	0 - 0
STD.5 CONC.-POS.	0 - 0
STD.6 CONC.-POS.	0 - 0
SD LIMIT	0.1
DUPLICATE LIMIT	400
SENSITIVITY LIMIT	0
ABS.LIMIT(INC/DEC)	0 - 0
PROZONE LIMIT	0 - 0
EXPECTED VALUE	35 - 80
UNIT	Mg/dl
INSTRUMENT FACTOR	1.00

HITACHI 704 پارامترهای کیت IRON برای دستگاه
با روش FERROZINE

TEST	FE
ASSAY CODE	2 POINT- 15 - 32
SAMPLE VOLUME (μl)	20
R 1 VOLUME (μl)	300 - 50 - NO
R 2 VOLUME (μl)	75- 20 - NO
WAVELENGTH (nm)	700 - 570
CALIBR. METHOD	1 - 0 - 0
STD.1 CONC.-POS.	0 - 1
STD.2 CONC.-POS.	* - *
STD.3 CONC.-POS.	0 - 0
STD.4 CONC.-POS.	0 - 0
STD.5 CONC.-POS.	0 - 0
STD.6 CONC.-POS.	0 - 0
SD LIMIT	0.1
DUPLICATE LIMIT	500
SENSITIVITY LIMIT	0
ABS.LIMIT(INC/DEC)	0 - 0
PROZONE LIMIT	0 - 0
EXPECTED VALUE	23 - 168
UNIT	μg/dl
INSTRUMENT FACTOR	1.00

HITACHI 704 پارامتر های کیت LDH برای دستگاه
با روش DGKC

TEST	LDH
ASSAY CODE	RATE A- 19 - 28
SAMPLE VOLUME (μl)	6
R 1 VOLUME (μl)	240 - 50 - NO
R 2 VOLUME (μl)	60 - 20 - NO
WAVELENGTH (nm)	405 - 340
CALIBR. METHOD	1 - 0 - 0
STD.1 CONC.-POS.	0 - 1
STD.2 CONC.-POS.	* - *
STD.3 CONC.-POS.	0 - 0
STD.4 CONC.-POS.	0 - 0
STD.5 CONC.-POS.	0 - 0
STD.6 CONC.-POS.	0 - 0
SD LIMIT	0.1
DUPLICATE LIMIT	200
SENSITIVITY LIMIT	0
ABS.LIMIT(INC/DEC)	6500 - 1
PROZONE LIMIT	0 - 0
EXPECTED VALUE	0 - 480
UNIT	U/L
INSTRUMENT FACTOR	1.00

HITACHI 704 پارامترهای کیت LDL برای دستگاه
با روش DIRECT

TEST	LDL
ASSAY CODE	2 POINT- 15 - 32
SAMPLE VOLUME (μl)	3
R 1 VOLUME (μl)	270 - 50 - NO
R 2 VOLUME (μl)	90 - 20 - NO
WAVELENGTH (nm)	700 - 600
CALIBR. METHOD	1 - 0 - 0
STD.1 CONC.-POS.	0 - 1
STD.2 CONC.-POS.	* - *
STD.3 CONC.-POS.	0 - 0
STD.4 CONC.-POS.	0 - 0
STD.5 CONC.-POS.	0 - 0
STD.6 CONC.-POS.	0 - 0
SD LIMIT	0.1
DUPLICATE LIMIT	400
SENSITIVITY LIMIT	0
ABS.LIMIT(INC/DEC)	0 - 0
PROZONE LIMIT	0 - 0
EXPECTED VALUE	0 - 130
UNIT	Mg/dl
INSTRUMENT FACTOR	1.00

HITACHI 704 پارامترهای کیت LIPASE برای دستگاه
با روش IFCC/DGKC

TEST	LIP
ASSAY CODE	RATE A - 22 - 28
SAMPLE VOLUME (μl)	6
R 1 VOLUME (μl)	250 - 50 - NO
R 2 VOLUME (μl)	50 - 20 - NO
WAVELENGTH (nm)	700 - 570
CALIBR. METHOD	1 - 0 - 0
STD.1 CONC.-POS.	0 - 1
STD.2 CONC.-POS.	* - *
STD.3 CONC.-POS.	0 - 0
STD.4 CONC.-POS.	0 - 0
STD.5 CONC.-POS.	0 - 0
STD.6 CONC.-POS.	0 - 0
SD LIMIT	0.1
DUPLICATE LIMIT	200
SENSITIVITY LIMIT	0
ABS.LIMIT(INC/DEC)	10000 - 0
PROZONE LIMIT	0 - 0
EXPECTED VALUE	0 - 60
UNIT	U/L
INSTRUMENT FACTOR	1.00

HITACHI 704 پارامتر های کیت MAGNESIUM برای دستگاه

TEST	MG
ASSAY CODE	1 POINT- 15 - 0
SAMPLE VOLUME (μl)	3
R 1 VOLUME (μl)	300 - 50 - NO
R 2 VOLUME (μl)	0 - 20 - NO
WAVELENGTH (nm)	700 - 546
CALIBR. METHOD	1 - 0 - 0
STD.1 CONC.-POS.	0.0 - 1
STD.2 CONC.-POS.	* - *
STD.3 CONC.-POS.	0 - 0
STD.4 CONC.-POS.	0 - 0
STD.5 CONC.-POS.	0 - 0
STD.6 CONC.-POS.	0 - 0
SD LIMIT	0.1
DUPLICATE LIMIT	500
SENSITIVITY LIMIT	0
ABS.LIMIT(INC/DEC)	0 - 0
PROZONE LIMIT	0 - 0
EXPECTED VALUE	1.8 - 2.6
UNIT	Mg/dl
INSTRUMENT FACTOR	1.00

پارامتر های کیت PHOSPHORUS برای دستگاه **HITACHI 704**
با روش UV

TEST	PHOS
ASSAY CODE	1 POINT - 15 - 0
SAMPLE VOLUME (μl)	3
R 1 VOLUME (μl)	300 - 50 - NO
R 2 VOLUME (μl)	0 - 20 - NO
WAVELENGTH (nm)	405 - 340
CALIBR. METHOD	1 - 0 - 0
STD.1 CONC.-POS.	0.0 - 1
STD.2 CONC.-POS.	* - *
STD.3 CONC.-POS.	0 - 0
STD.4 CONC.-POS.	0 - 0
STD.5 CONC.-POS.	0 - 0
STD.6 CONC.-POS.	0 - 0
SD LIMIT	0.1
DUPLICATE LIMIT	500
SENSITIVITY LIMIT	0
ABS.LIMIT(INC/DEC)	0 - 0
PROZONE LIMIT	0 - 0
EXPECTED VALUE	2.6 - 4.5
UNIT	Mg/dl
INSTRUMENT FACTOR	1.00

HITACHI 704 پارامترهای کیت TIBC برای دستگاه
با روش FERROZINE

TEST	TIBC
ASSAY CODE	2 POINT - 15 - 32
SAMPLE VOLUME (μl)	20
R 1 VOLUME (μl)	300 - 50 - NO
R 2 VOLUME (μl)	75 - 20 - NO
WAVELENGTH (nm)	700 - 570
CALIBR. METHOD	1 - 0 - 0
STD.1 CONC.-POS.	0 - 1
STD.2 CONC.-POS.	* - *
STD.3 CONC.-POS.	0 - 0
STD.4 CONC.-POS.	0 - 0
STD.5 CONC.-POS.	0 - 0
STD.6 CONC.-POS.	0 - 0
SD LIMIT	0.1
DUPLICATE LIMIT	500
SENSITIVITY LIMIT	0
ABS.LIMIT(INC/DEC)	0 - 0
PROZONE LIMIT	0 - 0
EXPECTED VALUE	120 - 470
UNIT	μg/dl
INSTRUMENT FACTOR	1.00

HITACHI 704 پارامترهای کیت TOTAL PROTEIN برای دستگاه
با روش BIURET

TEST	T.P
ASSAY CODE	1 POINT - 15 - 0
SAMPLE VOLUME (μl)	6
R 1 VOLUME (μl)	300 - 50 - NO
R 2 VOLUME (μl)	0 - 20 - NO
WAVELENGTH (nm)	700 - 546
CALIBR. METHOD	1 - 0 - 0
STD.1 CONC.-POS.	0.0 - 1
STD.2 CONC.-POS.	* - *
STD.3 CONC.-POS.	0 - 0
STD.4 CONC.-POS.	0 - 0
STD.5 CONC.-POS.	0 - 0
STD.6 CONC.-POS.	0 - 0
SD LIMIT	0.1
DUPLICATE LIMIT	400
SENSITIVITY LIMIT	0
ABS.LIMIT(INC/DEC)	0 - 0
PROZONE LIMIT	0 - 0
EXPECTED VALUE	6.5 - 8.8
UNIT	g/dl
INSTRUMENT FACTOR	1.00

HITACHI 704 پارامتر های کیت TRIGLYCERIDE برای دستگاه
با روش GOD

TEST	TG
ASSAY CODE	1 POINT - 32 - 0
SAMPLE VOLUME (μl)	3
R 1 VOLUME (μl)	300 - 50 - NO
R 2 VOLUME (μl)	0 - 20 - NO
WAVELENGTH (nm)	700 - 505
CALIBR. METHOD	1 - 0 - 0
STD.1 CONC.-POS.	0 - 1
STD.2 CONC.-POS.	* - *
STD.3 CONC.-POS.	0 - 0
STD.4 CONC.-POS.	0 - 0
STD.5 CONC.-POS.	0 - 0
STD.6 CONC.-POS.	0 - 0
SD LIMIT	0.1
DUPLICATE LIMIT	400
SENSITIVITY LIMIT	0
ABS.LIMIT(INC/DEC)	0 - 0
PROZONE LIMIT	0 - 0
EXPECTED VALUE	0 - 200
UNIT	Mg/dl
INSTRUMENT FACTOR	1.00

HITACHI 704 پارامتر های کیت URINE PROTEIN برای دستگاه

با روش GOD

TEST	TPU
ASSAY CODE	2 POINT - 15 - 32
SAMPLE VOLUME (μl)	20
R 1 VOLUME (μl)	280 - 50 - NO
R 2 VOLUME (μl)	70- 20 - NO
WAVELENGTH (nm)	0- 450
CALIBR. METHOD	1 - 0 - 0
STD.1 CONC.-POS.	0.0 - 1
STD.2 CONC.-POS.	100 - 2
STD.3 CONC.-POS.	0 - 0
STD.4 CONC.-POS.	0 - 0
STD.5 CONC.-POS.	0 - 0
STD.6 CONC.-POS.	0 - 0
SD LIMIT	0.1
DUPLICATE LIMIT	500
SENSITIVITY LIMIT	0
ABS.LIMIT(INC/DEC)	0 - 0
PROZONE LIMIT	0 - 0
EXPECTED VALUE	0 - 141
UNIIT	Mg/L
INSTRUMENT FACTOR	1.00

HITACHI 704 پارامتر های کیت UREA برای دستگاه
با روش UV

TEST	UREA
ASSAY CODE	RATE A - 18 - 21
SAMPLE VOLUME (μl)	3
R 1 VOLUME (μl)	240 - 50 - NO
R 2 VOLUME (μl)	60 - 20 - NO
WAVELENGTH (nm)	415 - 340
CALIBR. METHOD	1 - 0 - 0
STD.1 CONC.-POS.	0 - 1
STD.2 CONC.-POS.	* - *
STD.3 CONC.-POS.	0 - 0
STD.4 CONC.-POS.	0 - 0
STD.5 CONC.-POS.	0 - 0
STD.6 CONC.-POS.	0 - 0
SD LIMIT	0.1
DUPLICATE LIMIT	200
SENSITIVITY LIMIT	0
ABS.LIMIT(INC/DEC)	8000 - 1
PROZONE LIMIT	0 - 0
EXPECTED VALUE	17 - 43
UNIT	Mg/dl
INSTRUMENT FACTOR	1.00

HITACHI پارامترهای کیت URIC ACID Mono reagent برای دستگاه

704

با روش TOOS

TEST	UA
ASSAY CODE	1 POINT - 15 - 0
SAMPLE VOLUME (μl)	6
R 1 VOLUME (μl)	300 - 50 - NO
R 2 VOLUME (μl)	0 - 20 - NO
WAVELENGTH (nm)	700 - 546
CALIBR. METHOD	1 - 0 - 0
STD.1 CONC.-POS.	0.0 - 1
STD.2 CONC.-POS.	* - *
STD.3 CONC.-POS.	0 - 0
STD.4 CONC.-POS.	0 - 0
STD.5 CONC.-POS.	0 - 0
STD.6 CONC.-POS.	0 - 0
SD LIMIT	0.1
DUPLICATE LIMIT	400
SENSITIVITY LIMIT	0
ABS.LIMIT(INC/DEC)	0 - 0
PROZONE LIMIT	0 - 0
EXPECTED VALUE	2.3 - 8.2
UNIT	Mg/dl
INSTRUMENT FACTOR	1.00

HITACHI پارامتر های کیت URIC ACID Two reagent برای دستگاه

704

با روش TOOS

TEST	UA
ASSAY CODE	1 POINT- 32 - 0
SAMPLE VOLUME (μl)	6
R 1 VOLUME (μl)	150 - 50 - NO
R 2 VOLUME (μl)	150- 20 - NO
WAVELENGTH (nm)	700 - 546
CALIBR. METHOD	1 - 0 - 0
STD.1 CONC.-POS.	0.0 - 1
STD.2 CONC.-POS.	* - *
STD.3 CONC.-POS.	0 - 0
STD.4 CONC.-POS.	0 - 0
STD.5 CONC.-POS.	0 - 0
STD.6 CONC.-POS.	0 - 0
SD LIMIT	0.1
DUPLICATE LIMIT	400
SENSITIVITY LIMIT	0
ABS.LIMIT(INC/DEC)	0 - 0
PROZONE LIMIT	0 - 0
EXPECTED VALUE	2.3 - 8.2
UNIT	Mg/dl
INSTRUMENT FACTOR	1.00

HITACHI 704 پارامتر های کیت ZINC برای دستگاه

TEST	ZN
ASSAY CODE	1POINT -15 - 0
SAMPLE VOLUME (μl)	15
R 1 VOLUME (μl)	300 - 50 - NO
R 2 VOLUME (μl)	0 20 - NO
WAVELENGTH (nm)	700- 546
CALIBR. METHOD	1 0 - 0
STD.1 CONC.-POS.	0.0 - 1
STD.2 CONC.-POS.	* - *
STD.3 CONC.-POS.	0 - 0
STD.4 CONC.-POS.	0 - 0
STD.5 CONC.-POS.	0 - 0
STD.6 CONC.-POS.	0 - 0
UNIT	Micg/dl
SD LIMIT	0.1
DUPLICATE LIMIT	500
SENSITIVITY LIMIT	0
ABS.LIMIT(INC/DEC)	0 - 0
PROZONE LIMIT	0 - 0
EXPECTED VALUE	60 110
INSTRUMENT FACTOR	1.00

HITACHI 704 پارامترهای کیت CRP برای دستگاه
با روش LIA

TEST	CRP
ASSAY CODE	2 POINT - 15 - 32
SAMPLE VOLUME (μl)	3
R 1 VOLUME (μl)	240 - 50 - NO
R 2 VOLUME (μl)	60 - 20 - NO
WAVELENGTH (nm)	700 - 546
CALIBR. METHOD	1 - 0 - 0
STD.1 CONC.-POS.	0 - 1
STD.2 CONC.-POS.	* - *
STD.3 CONC.-POS.	0 - 0
STD.4 CONC.-POS.	0 - 0
STD.5 CONC.-POS.	0 - 0
STD.6 CONC.-POS.	0 - 0
SD LIMIT	0.1
DUPLICATE LIMIT	500
SENSITIVITY LIMIT	0
ABS.LIMIT(INC/DEC)	0 - 0
PROZONE LIMIT	0 - 0
EXPECTED VALUE	0 - 6
UNIT	Mg/L
INSTRUMENT FACTOR	1.00

