

Creatin kinase

Quantitative determination of creatin kinase (CK)

IVD For in-vitro diagnostic use only

2°C 8°C
Store at 2-8°C

INTENDED USE

For the quantitative determination of creatine kinase in human serum or plasma.


INTRODUCTION

Creatine kinase is a cellular enzyme with wide tissue distribution in the body. Its physiological role is associated with adenosine triphosphate (ATP) generation for contractile or transport systems. Elevated CK values are observed in diseases of skeletal muscle and after myocardial infarction^{1,2}.

Clinical diagnosis should not be made on a single test result; it should integrate clinical and other laboratory data.

PRINCIPLE OF THE METHOD

Creatine kinase (CK) catalyses the reversible transfer of a phosphate group from phosphocreatine to ADP. This reaction is coupled to those catalysed by hexokinase (HK) and glucose-6-phosphate dehydrogenase (G6P-DH):



The rate of NADPH formation, measured photometrically, is proportional to the catalytic concentration of CK present in the sample.

REAGENTS

R 1 Buffer	Imidazol pH 7.0	100 mmol/L
	Glucose	20 mmol/L
	Magnesium acetate	10 mmol/L
	EDTA	2 mmol/L
R 2 Substrate	ADP	2 mmol/L
	AMP	5 mmol/L
	di-Adenosine-5- pentaphosphate	10 mmol/L
	NADP ⁺	2 mmol/L
	Hexokinase (HK)	2500 U/L
	Glucose-6-phosphate dehydrogenase (G6P-DH)	1500 U/L
	N-acetyl cysteine	20 mmol/L
	Creatine phosphate	30 mmol/L

PREPARATION

Working reagent (WR):

Dissolve 1 tablet of R 2 Substrate with 2.5 mL of R 1.

Cap vial and mix gently to dissolve contents.

Stability: 5 days at 2-8°C or 24 hours at room temperature (15-25°C).

ADDITIONAL EQUIPMENT

4. Spectrophotometer or colorimeter measuring at 340 nm.
5. Thermostatic bath at 25°C/30°C/ 37°C (± 0.1°C).
6. Matched cuvettes 1.0 cm light path.
7. General laboratory equipment.

STORAGE AND STABILITY

8. All the components of the kit are stable until the expiration date on the label when stored tightly closed at 2-8°C, protected from light and contaminations prevented during their use.
9. Do not use the tablets if appears broken.
10. Do not use reagents over the expiration date.
11. Signs of reagent deterioration:

- Presence of particles and turbidity.
- Blank absorbance (A) at 340 nm ≥ 1.60 .

SAMPLES

Serum or plasma: Stability 7 days at 2-8°C, protected from light.

The creatine kinase activity decreases 10% after 1 day at 2-5°C or after

1 hour at 15-25°C.

PROCEDURE

1. Assay conditions:

Wavelength 340 nm

Cuvette 1 cm light path

Constant temperature 25°C / 30°C / 37°C

2. Adjust the instrument to zero with distilled water or air.

3. Pipette into a cuvette:

	25 - 30°C	37°C
WR (mL)	1.0	1.0
Sample (μL)	40	20

- Mix, incubate for 2 minutes.
- Read initial absorbance (A) of the sample, start the stopwatch and read absorbances at 1 minute intervals thereafter for 3 minutes.
- Calculate the difference between absorbances and the average
- absorbance differences per minute ($\Delta A/\text{min}$).

CALCULATIONS

25°- 30°C $\Delta A / \text{min} \times 4127 = \text{U/L CK}$

37°C $\Delta A / \text{min} \times 8095 = \text{U/L CK}$

Units: One international unit (IU) is the amount of enzyme that transforms 1 μmol of substrate per minute, in standard conditions. The concentration is expressed in units per litre of sample (U/L).

Temperature conversion factors To correct results to other temperatures multiply by:

Assay temperature	Conversion factor to		
	25°C	30°C	37°C
25°C	1.00	1.56	2.44
30°C	0.64	1.00	1.56
37°C	0.41	0.63	1.00

QUALITY CONTROL

Control sera are recommended to monitor the performance of assay procedures.

If control values are found outside the defined range, check the instrument, reagents and technique for problems.

Each laboratory should establish its own

Quality Control scheme and corrective

actions if controls do not meet the

acceptable tolerances. REFERENCE VALUES'

25°C 30°C 37°C

Men, up to 80 U/L 130 U/L 195 U/L

Women, up to 70 U/L 110 U/L 170 U/L

These values are for orientation purpose; each laboratory should establish its own reference range.

	Intra-assay (n=20)	
Mean U/L	166	450
SD	2.36	3.72
CV(%)	1.42	0.82

PERFORMANCE CHARACTERISTICS

Measuring range:

From detection limit of 1.35 U/L to linearity limit of 1000 U/L. If the results obtained were greater than linearity limit, dilute the sample 1/10 with NaCl 9 g/L and multiply the result by 10.

Precision:

Inter-Assay (n=20)	
165	446
2.26	5.17
1.37	1.16














Sensitivity:

1 U/L = 0.0001 AA/min.

Accuracy:

Results obtained using Atlas reagents (y) did not show systematic differences when compared with other commercial reagents (x).

The results of the performance characteristics depend on the analyzer used.

	Catalogue Number		Store at
	For In-Vitro Diagnostic use		Caution
	Number of tests in the pack		Read product insert before use
	Lot (batch) number		Manufacturer
	Fragile, handle with care		Expiry date
	Manufacturer fax number		Do not use if package is damaged
	Manufacturer telephone number		

INTERFERENCES

No interferences were observed with bilirubin up to < 20 mg/dL and hemoglobin up to 10 g/L. A list of drugs and other interfering substances with CK determination has been reported by Young.

REFERENCES

1. Abbot B et al. Creatinine kinase. Kaplan A et al. Clin Chem The C.V. Mosby Co. St Louis. Toronto. Princeton 1984: 1112-116.
2. Gerhardt W et al. Creatine kinase B-Subunit activity in serum after immunoinhibition of M-Subunit activity. Clin Chem 1979;(25/7): 1274-1280.
3. Young DS. Effects of drugs on Clinical Lab. Tests, 4th ed AACC Press, 1995.
4. Young DS. Effects of disease on Clinical Lab. Tests, 4th ed AACC 2001.
5. Burtis A et al. Tietz Textbook of Clinical Chemistry, 3rd ed AACC 1999.
6. Tietz N W et al. Clinical Guide to Laboratory Tests, 3rd ed AACC 1995.