

NOVIN BIO KIT

Total & Direct Bilirubin

DMSO. Colorimetric

IVD For in vitro diagnostic use only.

2°C  8°C Store at 2-8

INTENDED USE

For the determination of Total and Direct Bilirubin in human serum or plasma.

INTRODUCTION

Bilirubin is a breakdown product of hemoglobin, insoluble in water. It is transported from the spleen to the liver and excreted into bile. Hyperbilirubinemia results from the increase of bilirubin concentrations in plasma. Causes of hyperbilirubinemia:

Total bilirubin: Increase hemolysis, genetic errors, neonatal jaundice, ineffective erythropoiesis, and drugs. Direct bilirubin: Hepatic cholestasis, genetic errors, hepatocellular damage.

Clinical diagnosis should not be made on a single test result; it should integrate clinical and other laboratory data.

PRINCIPLE OF THE METHOD

Bilirubin is converted to colored azobilirubin by diazotized sulfanilic acid and measured photometrically. Of the two fractions presents in serum, bilirubin-glucuronide and free bilirubin loosely bound to albumin, only the former reacts directly in aqueous solution (bilirubin direct), while free bilirubin requires solubilization with dimethylsulphoxide (DMSO) to react (bilirubin indirect). In the determination of indirect bilirubin the direct is also determined, the results correspond to total bilirubin.

The intensity of the color formed is proportional to the bilirubin concentration in the sample.

REAGENTS

R1(Total Bilirubin)	Sulfanilic acid	30mmol/L
	Hydrochloric acid (HCl)	150mmol/L
R2(Direct Bilirubin)	Sulfanilic acid	30mmol/L
	Hydrochloric acid (HCl)	150mmol/L
	Dimethylsulphoxide (DMSO)	7 mol/L
R3	Sodium nitrite	29 mmol/L

EQUIPMENTS NEEDED BUT NOT PROVIDED

- Spectrophotometer or colorimeter measuring at 555 nm
- Matched cuvettes 1.0 cm light path.
- General laboratory equipment.

PRECAUTIONS

- Dimethylsulphoxide (DMSO): Irritant to eyes and skin. In case of contact with eyes, rinse immediately with plenty of water and seek medical advice.
- Hydrochloric acid (HCL): Irritant to eyes respiratory system and skin. In case of contact with eyes, rinse immediately with plenty of water and seek medical advice.

PREPARATION

- All the reagents are ready to use.

STORAGE AND STABILITY

- All the components of the kit are stable until the expiration date on the label when stored tightly closed at 2-8°C, protected from light and contaminations prevented during their use.
- Do not use reagents over the expiration date.
- **Signs of reagent deterioration:**
 - Presence of particles and turbidity.
 - Color development in R 2.

SAMPLES

Serum or plasma, free of hemolysis. Protect samples from direct light. Stability: Bilirubin is stable at 2-8°C for 4 days and 2 months at -20°C.

PROCEDURE

- Assay conditions:
Wavelength 555 nm (530-580)
Cuvette: 1 cm light path
Temperature 15-25°C
- Adjust the instrument to zero with distilled water.
Pipette into a cuvette:

TOTAL BILIRUBIN DETERMINATION:

	Blank	B. Total
R 1 (mL)		1000
R 3 (μL)	--	250
Sample	55	55

- Mix and incubate for exactly **5 minutes** at room temperature.
- Read the absorbance (A).

DIRECT BILIRUBIN DETERMINATION:

	Blank	B. Direct
R 2 (mL)		1000
R 3 (μL)	--	250
Sample (μL)	55	55

- Mix and incubate for exactly **5 minutes** at room temperature.
- Read the absorbance (A).

CALCULATIONS

1. TOTAL BILIRUBIN

(A) Sample - (A) Sample Blank x **19.1** = mg/dL bilirubin in the sample

Conversion factor: mg/dL x 17.1 = μmol/L.

2. DIRECT BILIRUBIN

(A) Sample - (A) Sample Blank x **14** = mg/dL bilirubin in the sample

Conversion factor: mg/dL x 17.1 = μmol/L

QUALITY CONTROL

If control values are found outside the defined range, check the instrument, reagents and calibrator for problems.

Each laboratory should establish its own Quality Control scheme and corrective actions if controls do not meet the acceptable tolerances.

REFERENCE VALUES

Bilirubin Total in adult Up to 1.10 mg/dL = Up to 18.81 μmol/L

Bilirubin Total in newborn <12mg/dL = <205.2 μmol/L

Bilirubin Direct Up to 0.25 mg/dL = Up to 4.27 μmol/L

These values are for orientation purpose; each laboratory should establish its own reference range.

PERFORMANCE CHARACTERISTICS

Bilirubin Total

Measuring range:

From *detection limit* of 0.00526 mg/L to *linearity limit* of 18 mg/dL.

If the results obtained were greater than linearity limit, dilute the sample 1/2 with NaCl 9 g/L and multiply the result by 2.

Precision:

	Intra-assay (n=20)		Inter-assay (n=20)	
Mean (mg/dL)	1.53	5.06	1.53	5.02
SD	0.03	0.05	0.03	0.11
CV (%)	1.73	1.01	1.92	2.18

Sensitivity:

1 mg/dL = 0.05074 A.

Bilirubin Direct

Measuring range:

From *detection limit* of 0.07 mg/L to *linearity limit* of 20 mg/dL.

If the results obtained were greater than linearity limit, dilute the sample 1/2 with NaCl 9 g/L and multiply the result by 2.

Precision:

	Intra-assay (n=20)		Inter-assay (n=20)	
Mean (mg/dL)	1.53	5.06	1.53	5.02
SD	0.03	0.05	0.03	0.11
CV (%)	1.73	1.01	1.92	2.18

Sensitivity:

1 mg/dL = 0.06856 A.

Accuracy:

Results obtained using reagents (y) did not show systematic differences when compared with other commercial reagents (x). The results obtained using 50 samples for Bilirubin D were the following:

Correlation coefficient (r)²: 0.96.

Regression equation: y= 0.71177x -0.05267.

The results obtained using 50 samples for Bilirubin T were the following:

Correlation coefficient (r)²: 0.991.

Regression equation: y= 0.82743x -0.0382.

The results of the performance characteristics depend on the analyzer used.

INTERFERENCES

Hemolysis causes decreased bilirubin values.

A list of drugs and other interfering substances with bilirubin determination has been reported.

NOTES

















1. For bilirubin determination in newborns, pipette 50 µL of sample. Multiply the result by 2.

REFERENCES

1. Kaplan A et al. Bilirubin. Clin Chem The C.V. Mosby Co. St Louis. Toronto. Princeton 1984; 1238-1241. 436 and 650.
2. Malloy H T. et al. The determination of bilirubin with the photoelectric colorimeter. J. Biol Chem 1937; 112, 2; 481-491.
3. Martinek R. Improved micro-method for determination of serum bilirubin. Clin Chim 1966: Acta 13: 61-170.
4. Young DS. Effects of drugs on Clinical Lab. Tests, 4th ed AACC Press, 1995.
5. Young DS. Effects of disease on Clinical Lab. Tests, 4th ed AACC 2001.
6. Burtis A et al. Tietz Textbook of Clinical Chemistry, 3rd ed AACC 1999.
7. Tietz N W et al. Clinical Guide to Laboratory Tests,

PPI1626A01

Rev A (02.09.2019)

	Catalogue Number		Temperature limit
	<i>In Vitro</i> diagnostic medical device		Caution
	Contains sufficient for <n> tests and Relative size		Consult instructions for use (IFU)
	Batch code		Manufacturer
	Fragile, handle with care		Use-by date
	Manufacturer fax number		Do not use if package is damaged
	Manufacturer telephone number		Date of Manufacture
	Keep away from sunlight		Keep dry